

# APT TESTING AND RESEARCH PRIVATE LIMITED

APPROVED BY FDA - MAHARASHTRA STATE LIC NO. - 37-PD/TL/7  
ISO 9001:2015 CERTIFIED LABORATORY (Regd. No. QMS.91.002.04.16)

Regd. Office: S. No. 36/1/1, M.N. 199, Vadgaon Khurd, Pune 411 041, Maharashtra, India.

• Tel.: +91 9607352020 • E-mail: info@ntc-apt.com, ntcpune@gmail.com

• Web: ntc-apt.com • CIN NO.: U74220PN2008PTC132043



PM509/2223/2048VIMa(po)

Date: 11/May/2022

## C E R T I F I C A T E

This is to certify that the Test item "PVC BLISTER FILM CLEAR", supplied by, SAROJ PACKAGING KHARUNI, P.O. MANPURA, BADDI-NALAGARH HIGHWAY DIST: SOLAN (H.P.) PIN: 174101; caused **NO TOXICITY** in Swiss albino mice, when test item extract was administered by oral route, according to the ISO: 10993-11:2017(E).

The test item was found to be "**NON TOXIC**" in Swiss albino mice by oral route.

The report of the toxicity test conducted has been submitted through Study Code No.

"PM509/2223/2048VIMa(po)".

This test report is valid for the specified test item tested at the time of study under the stated environmental conditions.

As the test item "PVC BLISTER FILM CLEAR", COMPLIES with the requirements for the Test for Systemic Toxicity by oral route in mice, it can be concluded that the test item "PVC BLISTER FILM CLEAR", is **BIOCOMPATIBLE** and SAFE FOR USE.

Hence it can be concluded the product is safe for use

ATR is approved by the Food & Drug Administration, Maharashtra State, Pune, through License No. 37-PD/TL/7.

Ms. M. D. Pandure, (M.Sc)

SD/DTM

Dr. K. G. Apte (M.Sc., Ph.D.)

TM/TFM